MODEL 6010T



Automated Thermometry Bridge

- DCC Technology
- Current Reversal
- Front and Rear Panel Inputs
- Accuracy: <0.05 PPM
- Noise: <2nV
- IEEE488 and Manual Operation

MODEL INFORMATION

The Model 6010T is designed for automated resistance thermometry applications and provides the best accuracy and convenience based on the most recent developments in current comparator technology. The Model 6010T is completely self-calibrated directly against two stable standard resistors of equal value for ratios of up to 1:1, and against national standards for ratios \leq 13:1.

The Model 6010T is a fully automated resistance ratio bridge based on the Direct-Current-Comparator (DCC) principle, avoiding all polarization effects in PRT's caused by AC bridges (NOTE 1). Using innovative technology, the 6010T's speed and measurement accuracy accounts for increased interest and preferred status among many national laboratories. Self-calibration for verifying the linearity of the DC Comparator can be performed at any time.

The Model 6010T is ideal for temperature measurement. The maximum ratio of 13:1

provides the ability to measure PRT's throughout their range utilizing only one standard resistor. Current reversal insures that all dc offsets and thermal emf's are eliminated. Reversals are selectable from 2 seconds and the 6010T will track temperature changes up to 10% of full scale.

The Model 6010T measures both ratio and resistance via rear and front panel connections and the number of inputs can be expanded to 40 when used with a ten and/or twenty channel, low thermal matrix scanner (Models 4210A/4220A). Measurements are performed automatically with Measurements International's 6010T operating software.

The 6010T includes a wide range of features specifically tailored for temperature metrology. These include programmable currents with V2 or 1/V2 excitation, selectable filters and manual or IEEE488 modes of operation.



Specifications:

•		
Accuracy:	<0.05 PPM	
Resolution:	0.001 PPM of Full Scale	
Measurement Time:	20 Seconds for Full Balance	
Variable Incremental Balance:	2 to 1000 Seconds	
Warm Up Time:	None	
Differential and Absolute Measurement:	Ratio and Ohms	
Thermometer Measurement Range:	0.1 to 1 k Ω (UUT and/or Standard)	
Traceability to National Standards:	Completely Self Checking w/ 2 Std. Resistors	s
Maximum Ratio:	13:1	
Linearity:	<0.01 PPM (Completely Self Checking)	
Bandwidth:	0.5 to 0.001 Hz (2 to 1000 Seconds)	
External Standards:	AC or DC Standard Resistors	
Sensor Current:	10μA to 150mA - v2 or 1/v2 at any value	
Bridge Balancing:	Front Panel or Software Controlled	
Lead Connections:	True Four Wire (No Series Lead Resistance)	
Analog Output:	Null Balance ±10V, Programmable 0-10V	
Stability:	<0.01 PPM/Year	
Filter Selection:	0.3s, 1.0s, 3s	
Sensor Current Accuracy:	100 PPM	
Noise:	<2nV	
Temperature Coefficient:	0	
Insulation Resistance:	>10 ¹¹ Typically 10 ¹²	
Output Impedance:	Infinite	
Ambient Temperature:	10°C to 35°C	
Dimensions (W x D x H): 266mm x 451mm x 306mm	Weight: 22.7 kgDisributed By:	
Operating Power: 50/60 Hz 40 VA 110/120/220/240 Vac		

How to Order:

Model: 6010T - Thermometry Bridge

NOTE 1:TEMPMEKO 99: CALIBRATION OF SPRTs IN THE SUB RANGE BETWEEN THE TRIPLE POINT OF Hg AND THE MELTING POINT OF Ga, Piero Marcarino, Peter P.M. Steur, Roberto Dematteis CNR - Istituto di Metrologia "G.Colonnetti" (IMGC), Torino, Italy

Form MI 66, Rev. 8, Dated 2012-03-28 (QAP19, App. "N")

Corporate Headquarters

Worldwide Offices

Measurements International PO Box 2359, 118 Commerce Drive Prescott, Ontario, Canada KOE 1TO

Phone: (613) 925-5934 Fax: (613) 925-1195 Email: sales@mintl.com **MI-USA** Phone: (407) 706-0328 Fax: (407) 706-0318 Email: sales@mintl.com

Toll Free: 1-866-684-6393

MI-China

Phone: 86-10-64459890 Fax: 86-10-64459871 Email: sales@mintl.com

MI-Europe

Phone: (420) 731-440-663 Fax: (420) 572-572-358 Email: sales@mintl.com



Data Subject to Change - Rev. 2

www.mintl.com

Toll Free: 1-800-324-4988

© Copyright 2010, 2012 Measurements International Limited. All rights reserved.